Iron VACUettes® Kit

K-6010D/R-6001D: 0 - 30 & 30 - 300 ppm K-6010A/R-6001A: 0 - 60 & 60 - 600 ppm **K-6010B/R-6001B**: 0 - 120 & 120 - 1200 ppm **K-6010C/R-6001C:** 0 - 1200 & 1200 - 12,000 ppm

K-6210D/R-6201D: 0 - 30 & 30 - 300 ppm

Soluble Iron (K-6010A, B, C, D) Ferrous Iron (K-6210D)

- 1. Fill the dilutor snapper cup to the -ml- mark with distilled water (fig. 1).
- 2. Fill the small micro-test tube approximately halfway with the sample to be tested (fig. 2).
- 3. Make sure that the VACUette tip is firmly attached to the ampoule tip.
- 4. Holding the VACUette almost horizontally. touch the tip to the contents of the micro-test tube (fig. 2).

NOTE: The capillary tip will fill completely with sample.

- 5. Required for R-6001D and R-6201D only: Pull the VACUette into a vertical position. A small portion of the collected sample should fall into the sleeve of the VACUette tip (fig. 3). NOTE: If none of the sample falls immediately, tap lightly on the shoulder of the ampoule.
- 6. Place the VACUette between the vertical tip quides on the inside of the dilutor snapper cup. Snap the ampoule tip. The ampoule will fill leaving a bubble for mixing (fig. 4).
- 7. To mix the ampoule, invert it several times. allowing the bubble to travel from end to end.
- 8. Dry the ampoule. Obtain a test result 1 minute after snapping the tip.



Figure 1




Figure 2




Figure 3



- 9. Obtain a test result using the appropriate comparator.
 - a. Low Range Comparator (fig. 5): Place the ampoule, flat end first, into the comparator. Hold the comparator up toward a source of light and view from the bottom. Rotate the comparator until the best color match is found.
 - b. High Range Comparator (fig. 6): Place the ampoule between the color standards until the best color match is found.



Figure 5



Total Iron Procedure K-6010A, B, C, D and K-6210D

1. Fill the large (5 mL) micro-test tube to the 2 mL mark with the sample to be tested.

- 2. Add A-6000 Activator Solution to the 3 mL mark. Cap the microtest tube and shake it to mix the contents. Wait 4 minutes.
- 3. After 4 minutes, shake the micro-test tube again, then perform the Soluble/Ferrous Iron Procedure using this pretreated sample.
- 4. **Multiply** test results by **1.5** for the correct **Total Iron** concentration.

Safety Information

Read SDS (available at www.chemetrics.com) before performing this test procedure. Wear safety glasses and protective gloves.

Test Method

The Iron VACUettes®1 test kit employs the phenanthroline chemistry. 2,3,4 Various metals will produce high test results. Some forms of insoluble iron (magnetite, ferrite, etc.) will show very low recoveries with this test.

- 1. VACUettes is a registered trademark of CHEMetrics, Inc. U.S. Patent Nos. 4,537,747 & 4,596,780 2. APHA Standard Methods, 23rd ed., Method 3500-Fe B - 1997
- 3. ASTM D 1068 77, Iron in Water, Test Method A
- 4. J.A. Tetlow and A.L. Wilson, "The Absorptiometric Determination of Iron in Boiler Feed-water, "Analyst. Vol. 89, p 442 (1964).

CHEMetrics, Inc., Midland, VA

www.chemetrics.com

May 19, Rev. 16